

**FlashDx Shenzhen Inc.**

**Address:** B6-4C, B5-A201, No.2533 Guanguang Road,  
 Fenghuang Community, Fenghuang Street, Guangming District  
 Shenzhen, Guangdong, P. R. China.  
**Phone:** +86 (0)755-86965752

## Certificate of Analysis

Report No. RQ250703001

<b>Product Name</b>	Respiratory Panel 1.1		
<b>Catalog No.</b>	RP-P10-10	<b>Specification</b>	10 tests/ box
<b>Lot No.</b>	S025F702		
<b>Manufacture Date</b>	2025-07-02	<b>Expiry Dated</b>	2026-07-01

**QC Test Results:**

<b>Test Document</b>	FlashDx Respiratory Panel 1.1 Test Quality Control Specification		
<b>Test Sampling Amount</b>	20 Tests		
<b>Test Item</b>	<b>Test Criteria</b>	<b>Test Result</b>	
Appearance	All components of the cartridge are complete and intact. The package label shall be clear and no wear. The aluminum foil bag containing the cartridge is sealed without damage.		Pass
Compliance rate of positive reference	The testing results of positive references shall be all positive for the corresponding assay tests.		Pass
Compliance rate of negative reference	The testing results of negative references shall be negative for all assay tests. Internal control is valid, $Ct \leq 35$		Pass
Limit of Detection (LoD)	The testing results of LoD reference shall be positive for corresponding assay tests.		Pass
<b>Overall Test Conclusion</b>		 <b>QC Pass</b> <b>QA Dept.</b>	

<b>Tested by:</b>  Ruina Li <b>Date:</b> 2025.07.03	<b>Reviewed by:</b>  Wanning Xian <b>Date:</b> 2025.07.03	<b>Approved by:</b>  Yongjun Yang <b>Date:</b> 2025.07.03
--	--	--