



FlashDx Shenzhen Inc.

Address: B6-4C, B5-A201, No.2533 Guanguang Road,  
Fenghuang Community, Fenghuang Street, Guangming District  
Shenzhen, Guangdong, P. R. China.  
Phone: +86 (0)755-86965752

## Certificate of Analysis

Report No. RQ2601080001

|                  |                               |               |               |
|------------------|-------------------------------|---------------|---------------|
| Product Name     | SARS-CoV-2 FluA FluB RSV Test |               |               |
| Catalog No.      | CoV2P4-10                     | Specification | 10 tests/ box |
| Lot No.          | S007G102                      |               |               |
| Manufacture Date | 2026-01-07                    | Expiry Date   | 2027-01-06    |

### QC Test Results:

|                                       |                                                                                                                                                                                 |             |
|---------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------|
| Test Document                         | FlashDx SARS-CoV-2 FluA FluB RSV Test Quality Control Specification                                                                                                             |             |
| Test Sampling Amount                  | 10 Tests                                                                                                                                                                        |             |
| Test Item                             | Test Criteria                                                                                                                                                                   | Test Result |
| Appearance                            | All components of the cartridge are complete and intact. The package label shall be clear and no wear. The aluminum foil bag containing the cartridge is sealed without damage. | Pass        |
| Compliance rate of positive reference | The testing results of positive references shall be all positive for the corresponding assay tests.                                                                             | Pass        |
| Compliance rate of negative reference | The testing results of negative references shall be negative for all assay tests. Internal control is valid, $C_t \leq 35$                                                      | Pass        |
| Limit of Detection (LoD)              | The testing results of LoD reference shall be positive for corresponding assay tests.                                                                                           | Pass        |
| Overall Test Conclusion               |                                                                                            |             |

|                                            |                                                  |                                                |
|--------------------------------------------|--------------------------------------------------|------------------------------------------------|
| Tested by:<br>Ruina Li<br>Date: 2026.01.08 | Reviewed by:<br>Wanming Xian<br>Date: 2026.01.08 | Approved by:<br>Yongjun Yang<br>Date: 2026.1.8 |
|--------------------------------------------|--------------------------------------------------|------------------------------------------------|