



FlashDx Shenzhen Inc.


Address: B6-4C, B5-A201, No.2533 Guanguang Road,
Fenghuang Community, Fenghuang Street, Guangming District
Shenzhen, Guangdong, P. R. China.
Phone: +86 (0)755-86965752

Certificate of Analysis

Report No. RQ260320001

Product Name	SARS-CoV-2 Test		
Catalog No.	CoV2-10	Specification	10 tests/ box
Lot No.	S010G304		
Manufacture Date	2026-03-20	Expiry Date	2027-03-19

QC Test Results:

Test Document	FlashDx SARS-CoV-2 Test Quality Control Specification	
Test Sampling Amount	10 Tests	
Test Item	Test Criteria	Test Result
Appearance	All components of the cartridge are complete and intact. The package label shall be clear and no wear. The aluminum foil bag containing the cartridge is sealed without damage.	Pass
Compliance rate of positive reference	The testing results of positive references shall be all positive for the corresponding assay tests.	Pass
Compliance rate of negative reference	The testing results of negative references shall be negative for all assay tests. Internal control is valid, $Ct \leq 35$	Pass
Limit of Detection (LoD)	The testing results of LoD reference shall be positive for corresponding assay tests.	Pass
Overall Test Conclusion		

Tested by: Ruina Li Date: 2026.03.20	Reviewed by: Wanming Xian Date: 2026.03.20	Approved by: Yongjun Yang Date: 2026.3.20
---	---	--