



FlashDx Shenzhen Inc.

Address: B6-4C, B5-A201, No.2533 Guanguang Road,  
Fenghuang Community, Fenghuang Street, Guangming District  
Shenzhen, Guangdong, P. R. China.


Phone: +86 (0)755-86965752

## Certificate of Analysis

Report No. RQ260330001

|                  |                               |               |               |
|------------------|-------------------------------|---------------|---------------|
| Product Name     | SARS-CoV-2 FluA FluB RSV Test |               |               |
| Catalog No.      | CoV2P4-10                     | Specification | 10 tests/ box |
| Lot No.          | S007G306                      |               |               |
| Manufacture Date | 2026-03-29                    | Expiry Date   | 2027-03-28    |

### QC Test Results:

|                                       |   |             |
|---------------------------------------|---|-------------|
| Test Document                         | FlashDx SARS-CoV-2 FluA FluB RSV Test Quality Control Specification   |             |
| Test Sampling Amount                  | 10 Tests  |             |
| Test Item                             | Test Criteria   | Test Result |
| Appearance                            | All components of the cartridge are complete and intact. The package label shall be clear and no wear. The aluminum foil bag containing the cartridge is sealed without damage. | Pass        |
| Compliance rate of positive reference | The testing results of positive references shall be all positive for the corresponding assay tests.   | Pass        |
| Compliance rate of negative reference | The testing results of negative references shall be negative for all assay tests. Internal control is valid, $Ct \leq 35$   | Pass        |
| Limit of Detection (LoD)              | The testing results of LoD reference shall be positive for corresponding assay tests.   | Pass        |
| Overall Test Conclusion               |    |             |

|   |   |  |
|---|---|--|
| Tested by:<br><i>Ruina Li</i><br>Date: 2026.03.30 | Reviewed by:<br><i>Wanning Xian</i><br>Date: 2026.03.30 | Approved by:<br><i>Yongjun Yang</i><br>Date: 2026.3.30 |
|---|---|--|